



CORRIGENDUM NO. 2
To the
CONTRACT NOTICE

Publication reference : MOH-GSD/RING/2022/29
Contract title: Supply of equipment and services for jenin Cardiac Center

According to site visit to the location of installation (Cardiac center-Jenin Hospital) which had been done on Monday 20/6/2022 please note the following notes regarding item no. (1) Cath Lab System:

1. MOH will delay the constructions of tiles for the floor and concrete bases for Cathlab table & bases for Floor mounted X-ray C-Arm after the contracting and it will be prepared by MOH with cooperation with the winner supplier.
2. MOH will delay the False ceiling works after the installation of Cath Lab and it will be prepared by MOH with cooperation with the winner supplier.
3. MOH will prepare the isolation (lead radiation shielding) of Cath Lab room walls & doors.
4. MOH will prepare vertical medical gases units (Vertical bed head units) qt. (2) in the treatment room and each unit will contain 2 Oxygen outlet , 2 medical air outlet ,2 vacuum outlet and 1 Nitrus outlet.
5. Lead glass for control room should be included by tenderers offers.
6. MOH will connect 3Phase cable with 300A contactor from main electrical cabinet to machines room of Cath Lab .
7. Supplier should supply and install electrical cabinet – in the machinery room- and its all needed accessories for which is special for Cath Lab.
8. Tenderers should separately price the following:
 - 3 Ton air conditioner qty(2)
 - 1.5 Ton air condition qty(3)
9. MOH will enlarge the window in the eastern side of the building of installation to enter the cath lab (to enter the boxes and system parts).
10. Supplier will sustain and repair any defect in the building may be happened during the entrance or installation of Cath Lab.

This CORRIGENDUM considered as a part of tender documents and will be taken in consideration during evaluating the tender.

Naser Al-Kahteeb

General Director of General Supplies Department

Chairperson of Central Tendering Committee